

Immunoassay










REF CMB1101 / CMB1102 / CMB1103 / CMB1104 / CMB1105



50 tests*1 / 100 tests*1 / 100 tests*2 / 100 tests*5 / 50 tests*2

NSE CLIA Microparticles

This assay is based on a chemiluminescent microparticle immunoassay (CLIA Microparticles) for the quantitative determination of NSE (Neuron Specific Enolase) in human serum.

All trademarks are properties of their respective owners.

Key to Graphical Symbols Used			
	batch code		use by
	manufacturer		contains sufficient for <n> tests
	<i>in vitro</i> diagnostic medical device		temperature limitation
	catalogue number		consult instructions for use
	authorized representative in the European Community		

	<p>OBELIS S.A. Bd. Général Wahis, 53 1030 Brussels Belgium</p> <p>AUTOBIO DIAGNOSTICS CO., LTD. No.87 Jingbei Yi Road National Eco & Tech Development Area Zhengzhou China 450016</p>
	



For any technical assistance please contact us in English at:
 Email: customerservice@autobio.com.cn
 Contact your local dealer for all product-related questions in your local language.

Introduction

The glycolytic enzyme enolase (2-phosph-D-glycerate hydrolyase) exists as several dimeric isoenzymes ($\alpha\alpha$, $\alpha\beta$, $\alpha\gamma$ and $\gamma\gamma$) composed of three distinct subunits, α , β , and γ . Three isoenzymes are found in human brain: $\alpha\alpha$, $\alpha\gamma$ and $\gamma\gamma$ ^[1]. The $\alpha\gamma$ and $\gamma\gamma$ -enolase isoenzymes are also known as neuron-specific enolase (NSE) as these isoenzymes initially were detected in neurons and neuronendocrine cells.

Elevated levels of NSE are commonly found in patients with malignant tumors with neuronendocrine differentiation, especially small cell lung cancer and neuroblastoma^[2]. Approximately 20% of the lung cancer is small cell lung cancer. Patients with small cell lung cancer show various proportions of $\alpha\gamma$ and $\gamma\gamma$ isoenzyme^[3].

NSE is reported to be useful diagnostic marker for lung cancer, neuroblastoma, melanoma, seminoma^[4] and in injury of central nervous system. In addition to the above, NSE can be a valuable tool in following-up the effect of chemotherapy of small cell lung cancer, in prognostic evaluation of patients with small cell lung cancer, and in differential diagnosis between cell lung cancer and non-small cell lung cancer.

Measurement Principle

This assay is based upon the two-step sandwich method. In the first step, the sample and anti-NSE coated microparticles are combined. During the incubation, NSE present in the sample binds to the antibodies coated microparticles. After the washing, in the second step, Enzyme Conjugate is added to the reaction mixture. During the incubation, a complex is generated among the microparticles, the NSE within the sample and enzyme-linked antibodies by immunological reactions. Chemiluminescent Substrate is added and the complex catalyzes substrate, resulting in a chemiluminescent reaction. The resulting chemiluminescent reaction is measured as RLU. The RLU is proportional to the amount of NSE in the sample.

Materials provided


1. Calibrators

6 vials lyophilized calibrator A through F. The matrix is phosphate buffer solution (PBS) containing BSA (bovine serum albumin). Contains a selection of preservatives.

Reconstitute each lyophilized calibrator with 1.0 mL of distilled water. Invert the calibrator several times to mix it completely. Then allow the reconstituted material to stand for at least 30 minutes.

2. Reagent pack

Reagent pack provided ready to use.

	50*1	100*1	100*2	100*5	50*2
Microparticles Solution	1.2mL*1	2.3mL*1	2.3mL*2	2.3mL*5	1.2mL*2
Enzyme Conjugate	5.5mL*1	11.0mL*1	11.0mL*2	11.0mL*5	5.5mL*2
Sample Diluent	3.0mL*1	5.5mL*1	5.5mL*2	5.5mL*5	3.0mL*2

● Microparticles Solution

NSE antibody coated microparticles in PBS buffer containing casein. Contains a selection of preservatives.

● Enzyme Conjugate

HRP (horseradish peroxidase) labeled mouse monoclonal anti-NSE in PBS buffer containing BSA (bovine serum albumin). Contains a selection of preservatives.

● Sample Diluent

BIS-TRIS and NaCl buffer. Contains a selection of preservatives.

Assay analyzers on which the kit can be used

- AutoLumo A2000 Plus
- AutoLumo A2000 Plus B
- AutoLumo A1000

The chemiluminescent microparticle immunoassay (CLIA Microparticles) is intended for use on Assay Analyzer which is AutoLumo A2000 Plus, AutoLumo A2000 Plus B or Autolumo A1000.

Materials Required but not Provided

1. Assay Analyzer
2. Reaction vessel(s) for sample and reagent reaction
3. Sample cups or tube(s) for sample containing
4. Diluent Universal
5. Chemiluminescent Substrate
6. System Wash for washing the pipetting needle
7. Wash Buffer used in the washing procedure
8. Distilled or deionized water

Metrological Traceability of Calibrators

The measurand or analyte in the NSE calibrators is traceable to the manufacturer's working calibrators. Traceability process is based on EN ISO 17511. The assigned values were established using representative samples from same lot of calibrator and are specific to the assay methodologies of the reagents. Values assigned by other methodologies may be different. Such differences, if present, may be caused by inter-method bias.

Warnings and Precautions

1. For professional use only.
2. Follow the instruction for use carefully. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this instruction for use.
3. Refer to the material safety data sheet and product labeling for any chemical hazards that may be present in this assay.
4. Handle the potentially contaminated materials and wastes safely according to local requirement.
5. This assay contains materials of animal origin. Bovine components originate from countries where bovine spongiform encephalopathy (BSE) has not been reported.
6. Do not smoke, drink, eat or use cosmetics in the working area.
7. Wear protective clothing and disposable gloves when dealing with samples and reagents. Wash hands after operations.
8. Conduct the assay away from bad ambient conditions. e. g. ambient air containing high concentration corrosive gas such as sodium hypochlorite acid, alkaline, acetaldehyde and so on, or containing dust.
9. Do not use reagents beyond the labeled expiry date.
10. Do not mix or use components from kits with different batch codes.
11. When storing the calibrators, be certain the vials are securely sealed.
12. Ensure the microparticles are resuspended before loading it on the analyzer.
13. Avoid foam formation in all reagents and sample types (samples, calibrators and controls).
14. Do not substitute any reagent in this kit from other manufacturers or other lots.
15. When any damage to the protective packaging or any change of analytical performance is observed, do not use the kit.

Storage

1. Store the kit at 2-8°C. Do not freeze. Avoid strong light. When stored as directed, all reagents are stable until the expiration date.
2. Refrigerate the reagent pack at 2-10°C for a minimum of 2 hours prior to use.
3. Store the unsealed reagents pack upright on the analyzer or 2-10°C for a maximum of 28 days. After 28 days, the reagent pack must be discarded. Once they are removed from the analyzer, store them at 2-10°C in an upright position.
4. Seal and return the remaining calibrators at 2-8°C immediately after the experiment, under which conditions the stability will be retained for 1 week, for longer use, store opened calibrators in aliquots and freeze at -20 °C for 1 month. Avoid multiple freeze-thaw cycles, do not freeze-thaw more than once.

Sample

1. Collect samples in accordance with correct medical practices.
2. Do not use heat-inactivated samples. Do not use sodium azide preservative in samples.
3. Do not use samples with obvious microbial contamination.
4. Sediments and suspended solids in samples may interfere with the test result which should be removed by centrifugation. Ensure that complete clot formation in samples has taken place prior to centrifugation. Some samples, especially those from patients receiving anticoagulant or thrombolytic therapy, may exhibit increased clotting time. If the sample is centrifuged before a complete clot forms, the presence of fibrin may cause erroneous results. Be sure that the samples are not decayed prior to use.
5. Prior to shipment, it is recommended that samples be removed from the clot or red blood cells.
6. Use caution when handling patient samples to prevent cross contamination. Use of disposable pipettes or pipette tips is recommended.
7. Insufficient processing of sample or disruption of the sample during transportation may cause depressed results.
8. Avoid grossly hemolytic, lipemic or turbid samples.
9. Sample shall be centrifuged within 1 hour after collection and to remove the serum, improper centrifugation or hemolytic samples may have the elevated concentrations of NSE. Cap and store the samples at 18-25 °C for no more than 6 hours, for longer use samples should be capped and stored at 2-8 °C up to 24 hours. Or freeze the samples that need to be stored or transported for more than 24 hours at -70°C. Avoid multiple freeze-thaw cycles. Mix thawed samples thoroughly by low speed vortex or by inverting 10 times. Visually inspect the samples, if layering or stratification is observed, continue mixing until samples are visibly homogeneous. After thawing, bring to room temperature and mix well by gently shaking.
10. Centrifuge the thawed samples containing red blood cells or particulate matter, or which are hazy or cloudy in appearance prior to use to ensure consistency in the results.
11. Notice that interfering levels of fibrin may be present in samples that do not have obvious or visible particulate matter.
12. If proper sample collection and preparation cannot be verified, or if samples have been disrupted due to transportation or sample handling, an additional centrifugation step is recommended. Centrifugation conditions should be sufficient to remove particulate matter.
13. For optimal results, inspect all samples for bubbles. Remove bubbles with a tip prior to analysis. Use a new tip for each sample to prevent cross contamination.

Measurement Procedure

1. Check the consumable materials

- Verify adequate volume of consumable materials is present prior to running the test.
- Refer to the Assay Analyzer's operation manual.

2. Load the kit

- Mix contents of new (unpunctured) reagent packs by gently inverting pack several times before loading on the analyzer. Avoid foam formation in all reagents. Don't invert the open (punctured) packs. If necessary, shake gently to mix horizontally after the first loading.
- Read the bar code on the reagent pack automatically to obtain the required parameters for the test.
- Analyzer can read the bar code on the reagent pack automatically to obtain the required parameters for the test.
- If the bar code cannot be read in exceptional cases, they can be recognized manually.
- Refer to the Assay Analyzer's operation manual.

3. Order tests

- Place the sample cups or tubes on the sample rack, 25 μ L of samples for each test. But consider the sample container and 150 μ L of system dead volumes, which can be refer to the appropriate Assay Analyzer manuals for the minimum sample volume required.
- Load the sample rack and input the sample information on the system software interface.
- Select "run" to start the test, the analyzer automatically conduct test operation. It performs the following functions:
 - Move the sample to the set point
 - Load a reaction vessel into the process path
 - Aspirate and transfer sample into the reaction vessel
 - Add Sample Diluent and Microparticles Solution to the reaction vessel
 - Mix, incubate and wash the reaction mixture
 - Add Enzyme Conjugate to the reaction vessel
 - Mix, incubate and wash the reaction mixture
 - Add Chemiluminescent Substrate
 - Measure chemiluminescent emission to determine the quantity of NSE in the sample
 - Discard the used reaction vessel
 - Calculate the result
- Refer to the Assay Analyzer's operation manual.

4. Calibrate the curve

- Analyzer can read the bar code on the reagent pack automatically to obtain the required parameters for the test.
- If the bar code cannot be read in exceptional cases, they can be recognized manually.
- Transfer the calibrators into the sample cups or tubes and place them on the sample rack. Conduct duplicate detection on the system.
- Load the sample rack and input calibrator information on the system software interface.
- Select "run" to start the test and generate the calibration curve, calibration is required every 28 days.
- Once a calibration curve is accepted and stored, all subsequent samples may be tested without further calibration unless:
 - Controls are out of range after repeated measurements
 - A reagent kit and Chemiluminescent Substrate with new batch code is used
 - Beyond the expiration date of a calibration curve
 - Important parts of the analyzer are replaced or repaired
- Refer to the Assay Analyzer's operation manual.

5. Dilute the sample

Samples with an NSE value exceeding 300ng/mL may be diluted via the program of the analyzer. Diluent Universal is used to dilute the samples. The software takes the dilution into account when reporting the result.

Measurement Results

The sample test results are determined automatically by the system software. The amount of NSE in the sample is determined from the measured light production by means of the stored calibration data. Refer to the Assay Analyzer's operation manual on reviewing sample results.

Control Procedure

The recommended control requirement for this assay is to purchase control materials separately and test them together with the samples within the same run. The result is valid if the control values fall within the acceptable ranges. When a control value is out of the specified range, it may indicate deterioration of the reagents or errors in technique. Associated test results may be invalid and may require retesting. Assay recalibration may be necessary. It is recommended that each laboratory establish its accepted range to ensure proper test performance.

Limitations of the Procedure

1. This assay is intended as an aid for the clinical diagnosis. Conduct this assay in conjunction with clinical examination, patient's medical history and other test results.
2. If the results are inconsistent with clinical evidence, additional testing is suggested to confirm the result.
3. Heterophilic antibodies and rheumatoid factors in samples may interfere with test results. Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with in vitro immunoassays. Patients routinely exposed to animals or animal serum products can be prone to this interference and anomalous values may be observed. Additional information may be required for diagnosis. This kind of samples is not suitable to be tested by this assay.
4. Performance of this test has not been established with neonatal samples.
5. Patients who have received mouse monoclonal antibodies for either diagnosis or therapy can develop HAMA (human Anti-mouse antibodies). HAMA can produce either falsely high or falsely low values in immunoassays which use mouse monoclonal antibodies. Additional information may be required for diagnosis.
6. This test measures concentrations within the range of 0.5-300 ng/mL. If NSE concentrations above the measuring range to be expected, it is recommended to dilute samples with Diluent Universal, the recommended dilution is 1:3 of this test, under this condition, allowing samples to be up to approximately 1200 ng/mL.

Biological Reference Interval

By testing serum samples from 375 individuals defined as normal by a clinician, results are as follow:

20 ng/mL (95th percentile)

17.89-20.45 ng/mL (95% confidence interval of the percentile)

It is recommended that each laboratory establish its own normal range which may be unique to the population it serves depending upon geographical, patient, dietary, or environmental factors.

Performance Characteristics

1. Measurement Precision

3 samples were assayed in duplicate, twice per day across 20 testing days. Data from this study are summarized in the following table.

Sample	n	Mean (ng/mL)	Within-run	Total
			%CV	%CV
1	80	22.19	2.82	5.04
2	80	42.87	7.18	8.82
3	80	106.22	3.70	5.58

*Representative data; results in individual laboratories may vary from these data.

2. Analytical Sensitivity

Limit of Blank: 0.5 ng/mL

Limit of Detection: 1.0 ng/mL

Limit of Quantitation: 3.0 ng/mL with a coefficient of variation of $\leq 20\%$

3. Analytical Specificity

No interference with 100 mg/dL of bilirubin, 1500 mg/dL of triglyceride.

4. Measurement Accuracy by Correlation

A study was performed where samples were tested using this assay and an NSE Test which was already available on the market. Data were analyzed and are summarized in the following table.

Correlation Method	Number of Samples	Intercept	Slope	Correlation Coefficient
Linear Regression	296	2.3705	0.9571	0.9762

Literature References

1. Multifunctional α -enolase: its role in diseases. Cell Mol Life Sci. 58 (7): 902–20.
2. Molecular structure of the human muscle-specific enolase gene (ENO3). Biochem J. 275 (Pt 2): 427–33.
3. Larsen, JE; Minna D. Molecular biology of lung cancer: clinical implications. Clinics in Chest Medicine 32 (4): 703–740. 2011.
4. Wibe E., Paus E., Aamdal S., Neuron Specific Enolase (NSE) in serum of Patients with Malignant Melanoma, Cancer Letters, 52:29-31,1990.